



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204112	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/26/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Brookfield Police Department, 116 W. Brooks, Brookfield, MO		TIME OF INSPECTION 11:43 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02/26/14 1143
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13001 EXP. DATE 03/07/2015	
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G 6815 EXP. DATE 04/03/2014	
<input checked="" type="checkbox"/> CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	

TEST 1 ➡ .099	TEST 2 ➡ .099	TEST 3 ➡ .099
---------------	---------------	---------------

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

Meets Department of Health Standards

INSPECTING OFFICER

SIGNATURE <i>Lane P. Quinn</i>	PRINT FULL NAME Lane P. Quinn
TYPE II PERMIT NUMBER/EXPIRATION DATE 240025 / 01-28-16	TELEPHONE NUMBER (660) 385-2132

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
LANE P QUINN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/28/2014

NUMBER 240025

EXPIRES 1/28/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

'80-0771 (6-10)

LAB-4 (RS-10)

	STATE OF MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES BREATH ALCOHOL PROGRAM
	INSTRUMENT OPERATOR CARD
<small>The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.</small>	
Operator QUINN, LANE Permit No 240025 Date Issued 1/28/2014 Date Expires 1/28/2016	

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISQUON STATE HIGHWAY DEPTAL

ONE UNIVERSITY AVE. RUMBLE 20412

06/26/18

ISSUED BY: J. L. WALKER

SUBJECT: J. L. WALKER

DATE: 06/26/18

TIME: 11:00 AM

LOCATION: UNIVERSITY AVE. RUMBLE

REASON: J. L. WALKER

ISSUED BY: J. L. WALKER

ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00

Signature

TPR J. L. Walker

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISQUON STATE HIGHWAY DEPTAL

ONE UNIVERSITY AVE. RUMBLE 20412

06/26/18

ISSUED BY: J. L. WALKER

SUBJECT: J. L. WALKER

DATE: 06/26/18

TIME: 11:00 AM

LOCATION: UNIVERSITY AVE. RUMBLE

REASON: J. L. WALKER

ISSUED BY:

ISSUED BY:

ISSUED BY:

ISSUED BY:

ISSUED BY:

ISSUED BY:

ISSUED BY:

ISSUED BY:

ISSUED BY:

ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00
ISSUED BY:	06/26	11:00

Operator Signature

TPR J. L. Walker

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

PLEASE PRINT NAME, ADDRESS, PHONE
AND INTERVIEW DATE ON REVERSE END OF
SLIP ONLY
THANKS

DATE OF INTERVIEW

INTERVIEWER

INTERVIEW LOCATION

INTERVIEW TIME

INTERVIEW DATE

INTERVIEWER

INTERVIEW LOCATION

INTERVIEW TIME

INTERVIEW DATE

INTERVIEWER

INTERVIEW LOCATION

INTERVIEW TIME

INTERVIEW DATE

INTERVIEWER

INTERVIEW LOCATION

INTERVIEW TIME

Operator Signature

TPR L.P. Quinn